

PRODUCT DESCRIPTION						
Brand Name	Virta Classic Nitrile Powder Free, Non Sterile Examination Gloves.					
Intended Use	A patient examination glove is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between patient and examiner					
Product Conformance	ASTM D6319, EN 455 part 1 & 2					
Material	Nitrile (Carboxylated Acrylonitrile Butadiene)					
Cuff Finishing	Beaded					
Weight	3.5g					
Color	Blue					
Type	Ambidexterous					
Surface Finish	Finger Textured					
SPECIFICATIONS						
Dimension	Size	Palm Width	Length (mm)			
	XS	75 ± 5	≥240			
	S	85 ± 5	≥240			
	M	95 ± 5	≥240			
	L	105 ± 5	≥240			
	XL	115 ± 5	≥240			
Thickness		Single Wall (mm) Specification Min	Single Wall (mm) Specification Max			
	Palm	0.06	0.09			
	Finger	0.08	0.14			
	Cuff	0.05	0.08			
Physical Properties		Before Aging		After Aging		
	Tensile Strength (MPa)	14		14		
	Ultimate Elongation (%)	500		400		
Residual Powder	2.0mg / glove maximum					
Quality Specification	Inspection Criteria			Sampling Plan/AQL		
	Dimension			S2 – AQL 4.0		
	Physical Properties			S2 – AQL 4.0		
	Freedom from Holes			GI – 1.0		
	Visual Defects – Major			GI – 2.5		
	Visual Defects – Minor			GI – 4.0		
PACKAGE CONFIGURATION AND LABELLING						
Packing	100 gloves (by weight) per inner dispenser. 10 dispensers per outer carton					
Labelling	Dispensers and outer cartons shall have the following information: <ul style="list-style-type: none"> • Product description including size and quantity of gloves • Country of origin of Manufacturer • Single use only symbol, and non-sterile • Lot Number and date of manufacture 					

Product Certifications

ASTM: 6319, EN 455 part 1 & 2

02.08.2023, VCSPC_1.0