

PRODUCT DESCRIPTION					
Brand Name	Virta Lite Nitrile Powder Free, Non Sterile Examination Gloves. Hypoallergenic				
Intended Use	A patient examination glove is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between patient and examiner				
Product Conformance	ASTM D6319, EN 455 part 1 & 2				
Material	Nitrile (Carboxylated Acrylonitrile Butadiene)				
Cuff Finishing	Beaded				
Weight	2.9g				
Color	Indigo				
Type	Ambidexterous				
Surface Finish	Finger Textured				
SPECIFICATIONS					
Dimension	Size	Palm Width	Length (mm)		
	XS	75 ± 5	≥240		
	S	85 ± 5	≥240		
	M	95 ± 5	≥240		
	L	106 ± 5	≥240		
	XL	115 ± 5	≥240		
Thickness		Single Wall (mm) Specification	Typical Value		
	Palm	≥0.05	0.06		
	Finger	≥0.08	0.09		
Physical Properties		Un-Aged Specification	Typical Value	Aged Specification	Typical Value
	Tensile Strength (MPa)	≥14	24	≥14	36
	Ultimate Elongation (%)	≥500	580	≥400	600
Residual Powder	2.0mg / glove maximum				
Quality Specification	Inspection Criteria		Sampling Plan/AQL		
	Dimension		S2 – AQL 4.0		
	Physical Properties		S2 – AQL 4.0		
	Water Tight Test		G1 – AQL 1.5		
	Visual Inspection – Major		G1 – AQL 2.5		
	Visual Inspection – Minor		G1 – AQL 4.0		
	Residual Powder		n = 5 (average)		
	Packaging Labelling		n = 5 (accept, no defects)		
Glove Physical Counting		n = 5 (accept, no defects)			
PACKAGE CONFIGURATION AND LABELLING					
Packing	200 gloves (by weight) per inner dispenser. 10 dispensers per outer carton				
Labelling	Dispensers and outer cartons shall have the following information: <ul style="list-style-type: none"> • Product description including size and quantity of gloves • Country of origin of Manufacturer • Single use only symbol, and non-sterile • Lot Number and date of manufacture 				

Product Certifications

ASTM: 6319, EN 455 part 1 & 2

09.20.2023, VLSPC_5.0