



| PRODUCT DESCRIPTION | | | | | |
|-------------------------------------|--|---|---|--------------------|---------------|
| Brand Name | Virta Soothe Nitrile Powder Free, Non Sterile Examination Gloves. | | | | |
| Intended Use | A patient examination glove is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between patient and examiner. Accelerator Free (Low Derma Potential), Tested against Chemotherapy Drugs and Concentration | | | | |
| Product Conformance | ASTM D6319, EN 455 part 1 & 2, D6978 | | | | |
| Material | Nitrile (Carboxylated Acrylonitrile Butadiene) | | | | |
| Cuff Finishing | Beaded | | | | |
| Weight | 3.5g | | | | |
| Color | Green | | | | |
| Type | Ambidexterous | | | | |
| Surface Finish | Finger Textured | | | | |
| SPECIFICATIONS | | | | | |
| Dimension | Size | Palm Width | Length (mm) | | |
| | XS | 75 ± 5 | ≥240 | | |
| | S | 85 ± 5 | ≥240 | | |
| | M | 95 ± 5 | ≥240 | | |
| | L | 106 ± 5 | ≥240 | | |
| Thickness | | Single Wall (mm) Specification Min | Single Wall (mm) Specification Max | | |
| | Cuff | ≥0.04 | 0.09 | | |
| | Palm | ≥0.05 | 0.06 | | |
| | Finger | ≥0.08 | 0.10 | | |
| | Beading | ≥0.50 | 0.60 | | |
| Physical Properties | | Before Aging | | After Aging | |
| | | Specification | Typical Value | Specification | Typical Value |
| | Tensile Strength (MPa) | 14min | 24 | 14min | 28 |
| | Ultimate Elongation (%) | 500min | 550 | 400min | 570 |
| Residual Powder | 2.0mg / glove maximum | | | | |
| Quality Specification | Inspection Criteria | Sampling Plan/AQL | | | |
| | Dimension | S2 – AQL 4.0 | | | |
| | Physical Properties | S2 – AQL 4.0 | | | |
| | Water Tight Test | G1 – AQL 1.0 (Single Normal Sampling) | | | |
| | Freedom from Holes | GI – 1.0 | | | |
| | Visual Inspection – Major | GI – 2.5 | | | |
| | Visual Inspection – Minor | GI – 4.0 | | | |
| | Packaging Labeling | n = 5 (AND) | | | |
| Glove Physical Counting | n = 5 (AND) | | | | |
| PACKAGE CONFIGURATION AND LABELLING | | | | | |
| Packing | 100 gloves (by weight) per inner dispenser. 10 dispensers per outer carton | | | | |
| Labelling | Dispensers and outer cartons shall have the following information: <ul style="list-style-type: none"> • Product description including size and quantity of gloves • Country of origin of Manufacturer • Single use only symbol, and non-sterile • Lot Number and date of manufacture | | | | |